



- A Magnet
- B Positioning Coil
- C Position Sensing Device
- D Gas Flow inlet connection from Tank (from the Interconnection Block)
- E Variable Orifice for gas flow control to the Patient Breathing Circuit
- F dP Servo Differential Pressure Sensor connectors
- G Outlet connection to the Safety Valve Block