

SLE 5000 INFANT VENTILATOR

Mode Select **SIMV** Standby

PSV Setup Support OFF

Apnoea Setup Disabled when backup rate is 20 BPM or higher.

TTV ON

Additional pressure support can be given.

Ti Meas. 0.60

BPM Tot 20

Trigger 0

Vte (ml) 13.28

Vmin (l) 0.282

Leak (%) 5

Resist 351.00

Compl. 0.60

C20/C 1.90

Mean P 6

No Current Alarms 0 0

Hist Adjust

BPM 20

Max Ti 1.40

0 Sec. 3

Manual Breath

PEEP 2

0 mbar 20

Max PIP 24

0 mbar 65

Tidal Vol 14

3 ml 200

FiO2 21

21 % 100

PROXIMAL AIRWAY

EXHALATION PORT FROM PATIENT

FLOW SENSOR

FRESH GAS TO PATIENT